Shoulder Cryo/Cuff[™] Gravity Patient Information

_

What is the Cryo/Cuff Gravity?

The Aircast Cryo/Cuff Gravity combines cold and focal compression to better manage pain and reduce swelling. Continuous cold and focal compression are delivered using a simple, gravity-fed delivery system.

How is the Cryo/Cuff Gravity used?

STEP 1 - Prepare cooler (Figure 1)


- 1. Connect the blue tube to cooler.
- 2. Add cold water to line inside cooler.
- 3. Fill with ice to the line inside the cooler.
- 4. Lay insulation disk on top of ice. Attach cooler lid snugly.
- 5. Allow five minutes with occasional shaking to chill water.


STEP 2 - Apply EMPTY Shoulder Cryo/Cuff (Figure 2)

- Use only Aircast Cryo/Cuffs with Aircast Cryo/Cuff Gravity.
- Snugly secure strap around torso in front and back.
 Adjust underarm strap (first time only). Secure front and back elastic straps.
- Carefully and completely read instructions supplied with the specific Shoulder Cryo/Cuff prior to application.
 Please follow the instructions below to continue cooler preparation.

STEP 3 - Fill and pressurize Cryo/Cuff (Figure 3)

- Connect blue tube to Cryo/Cuff.
- Open cooler air vent.
- Raise cooler above Cryo/Cuff until cuff is full –
 NOTE: To <u>AVOID</u> excessive pressure during use,
 <u>DO NOT</u> raise the cooler higher than 15" (38 cm)
 above the Cryo/Cuff.
- Close cooler air vent.
- Cooler can be disconnected by pressing quick-disconnect.







To re-fill cooler:

Rechill Cryo/Cuff as needed.

IMPORTANT: <u>EMPTY</u> the Cryo/Cuff <u>COMPLETELY</u> while still attached to the cooler.

When Cryo/Cuff is empty then empty the cooler. Repeat steps 1-5 in **STEP 1** to prepare the cooler for refilling.

To clean cooler:

The cooler can be cleaned with disinfectant wipes. The tube assembly can be cleaned with common mild detergent and must be completely submerged and hand dried.

To troubleshoot:

- Make sure unit has ice and water filled to the line indicated inside the cooler.
- 2. Make sure to fill cooler and Cryo/Cuff as instructions indicate.
- 3. Check Cryo/Cuff connection; make sure to listen for a click when connecting the hose to the Cryo/Cuff.
- Make sure the Cryo/Cuff is completely <u>EMPTY</u> before refilling.

↑ CAUTION

- DO NOT use an elastic wrap with the Cryo/Cuff Gravity.
- Dressing used under the Cryo/Cuff should be applied lightly.
- Reduce pressure with any sense of discomfort, numbness or tingling of the limb.
- EMPTY the Cryo/Cuff after each use, and prior to refilling.
- Patient skin condition checks must occur every 1-2 hours on a routine basis.



FOR PRODUCT SUPPORT CALL 1-888-405-3251



DIOTIC